



Product Service

EC-Attestation of Conformity

No. N8 08 04 66679 001

Holder of Certificate: Unitma Co., Ltd.

3F, Chungmyeong B/D, 224-8, Jamsil-Dong
Songpa-Ku, Seoul 138-220
REPUBLIC OF KOREA

Product:

**Electrical equ. for measurement, control
and laboratory use
(Automated Tissue Microarrayer)**

This EC-Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2006/95/EC relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.:

CPSA0114148

Date, 2008-04-23

After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 2



Product Service

EC-Attestation of Conformity
No. N8 08 04 66679 001

Model(s): UATM-272

Parameters:

Rated input voltage:	100-120 V a.c., 200-240 V a.c.
Rated frequency:	50/60 Hz
Rated input power:	3.0 A, 1.5 A
Protection class:	I
Degree of protection against ingress of liquids:	IPX0

Tested according to:

- EN 61010-1:2001
- EN 61010-2-081:2002
- EN 61010-2-101:2002



Product Service

EC-Attestation of Compliance

No. E8N 08 04 66679 002

Holder of Certificate: Unitma Co., Ltd.
 3F, Chungmyeong B/D, 224-8, Jamsil-Dong
 Songpa-Ku, Seoul 138-220
 REPUBLIC OF KOREA

Name of Object: Laboratory Equipment
 Automated Tissue Microarrayer

Model(s): UATM-272

Description of Object: Rated input voltage: 100-120 / 200-240 V a.c.
 Rated input frequency: 50/60 Hz

Tested according to: EN 61326:1997+A1:1998+A2:2001+A3:2003 Class A

This EC-Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

Test report no.: EMC-CE-1213

Date, 2008-04-24



CE After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.