

Certificate of Compliance

Certificate **080510-E318534**
Report Reference **E318534, 2008 May 10**
Issue Date **2008 May 10**

Page 1 of 1



Issued to: **Unitma Co Ltd**

**3RD FL, CHUNGMYEONG B/D 224-8
JAMSIL-DONG SONGPA-KU SEOUL 138-220 KOREA**

*This is to certify that
representative samples of*

**Laboratory Use Electrical Equipment
AUTOMATED TISSUE MICROARRAYER, Model UATM-272**


*Have been investigated by Underwriters Laboratories Inc.® in
accordance with the Standard(s) indicated on this Certificate.*

Standard(s) for Safety:

**UL 61010-1, 2nd Edition, 2005-07-22 (Electrical Equipment for Measurement, Control,
and Laboratory Use; Part 1: General Requirements)
CAN/CSA-C22.2 No. 61010-1, 2nd Edition, 2004-07, (Electrical Equipment for
Measurement, Control, and Laboratory Use; Part 1: General Requirements)**

Additional Information: **N/A**

Only those products bearing the UL Listing Mark for the US and Canada should be considered as being covered by UL's Listing and Follow-Up Service meeting the appropriate requirements for US and Canada.

The UL Listing Mark for the US and Canada generally includes: the UL in a circle symbol with "C" and "US" identifiers:  the word "LISTED"; a control number (may be alphanumeric) assigned by UL; and the product category name (product identifier) as indicated in the appropriate UL Directory.

Look for the UL Listing Mark on the product

Issued by: **Sunhwa Shim**

Reviewed by: **Chanyo Won**

Sunhwa Shim, Engineering Project Handler

Chanyo Won, Staff Engineer

UL Korea Ltd.

UL Korea Ltd.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.
For questions in Korea you may call +82 2009 9000.